

- Copy
Quotation by Email

10066 - 590 -
NO. PUR / Store / GMC / Akola / /18
Office of the Dean Govt. Medical College
& Hospital Akola -44-1001 Date : 15 / 12 / 2018

TO,


Subject :- Submission of Quotation for the Supply Endo Master with Apex Locator
to Dean Govt. Medical College, Akola .

Dear Sir,

You are requested to submit your sealed quotation for the supply of
Machine/Surgical Item to this office with the terms and conditions mentioned
below

- 1) The quotation should be sent in sealed cover (duly sealed wax) by the post
or by hand delivery so as to reach this office on or before 26 / 12 / 2018
- 2) Quotation will be opened in the office of the dean on 27 / 12 / 2018
- 3) Quotations which are received late will not be accepted under any circumstances.
- 4) The words quotation for and last date, must be mentioned on the envelope.
- 5) Rate should be for free delivery at the Dean, Govt. Medical College, Akola,
premises only.
- 6) Quoted goods should be strictly according to the make & the specifications asked in
enquiry, and the make of the item by you should be stated in your quotation.
- 7) In case orders are placed with you, the order will have to be executed in full
within the stipulated time. If you fail to comply, the orders may be cancelled
and you will be declared a defaulter, no enquiry will be sent in the future without
assigning any reason.
- 8) The undersigned reserves the right to accept or reject any or all quotations. The
quotation should be sent in the name of the / Dean, medical college & Hospital,
Akola.
- 9) Please quote the sales tax / GST No. registration no. in your quotation

Sr.no	Name of Reagent	Qty
1	Endo Master with Apex Locator	
	(As per Specification)	


Dean
Govt. Medical College &
Hospital, Akola

Govt. Medical College, Akola.
Dept. of Dentistry

Specifications for Endomotor with Apex Locator

Sr.No.	Name of the Item	Specifications
1.	Endo Master (Endo motor with Apex locator)	<p>Apex locator with integrated Micromotor or Assorted apex locator and Endomotor for mechanical root canal preparation.</p> <p>Endomotor with Apex Locator: It should contain</p> <ul style="list-style-type: none"> - Main Unit - Motor Hand piece brushless - Testing Cord - Users Manual - Apex Locator: Included and working simultaneously and coordinated with motor - Contra Angle 1:1 Included <p>Specifications</p> <ul style="list-style-type: none"> - Voltage DC8.4V - Power 5W - Torque 6-52mn - Display Reflecting LCD Screen - Battery Chargeable Li-on Battery 3.7 x 2 - Current < 500mA <p>Features</p> <ul style="list-style-type: none"> - Adjustable Speed - 22 Settings for Adjustable Ranging from 100-625 rpm - Adjustable Torque - The Torque exceeds the regulated value, the file will reverse to avoid break-off. Torque control setting from 1 to 4 Nos. Automatic stop (torque/apex), - Saving Function - Nine groups of program data can be saved - Easy operating with LCD Screen - Built in Rechargeable Battery - Start the rotation of the ni-ti files with foot control pedal <p>Contra angle with microhead for Ni Ti files</p> <ul style="list-style-type: none"> - 1 year warranty <p><u>Assorted Endomotor:</u></p> <p>FEATURES</p> <ul style="list-style-type: none"> • Thin ergonomic design for better access to hard to reach areas