

- Copy
Quotation by Email

10068 - 589 -
NO. PUR / Store / GMC /Akola/ /18
Office of the Dean Govt. Medical College
& Hospital Akola 444001 Date : 15 /12 /2018

TO,


Subject :- Submission of Quotation for the Supply Digital Audiometer Set
to Dean Govt. Medical College, Akola .

Dear Sir,

You are requested to submit your scaled quotation for the supply of
Machine/Surgical Item to this office with the terms and conditions mentioned
below

- 1)The quotation should be send in scaled cover (duly sealed wax) by the post
or by hand delivery so as to reach this office on be 26 /12/ 2018
- 2) Quotation will be open in the office of the dean on 27 /12/ 2018
- 3)Quotation which are received late will not be accepted and any circumstances.
- 4) The words quotation for and last date , must be mentioned on envelope .
- 5) Rate should be for free delivery at the Dean , Govt. Medical college , Akola.
premises only .
- 6)Quoted good should be strictly according to the make & the specificationasked in
enquiry, and make of item by you should be stated in your quotation .
- 7)In case the order are placed with you , the order will have to be executed in full
within the stipulated time . If you fail to comply , the orders may be cancelled
and you will be declared defaulter , no enquiry will be sent in future without
assigning any reason.
- 8) The undersigned reserves the right to accept or reject any or all quotation. . The
quotation should be sent in the name of the / Dean , medical college & Hospital,
Akola.
- 9) Please quote the sale tax /GST No. registration no . in your quotation

Sr.no	Name of Reagent	Qty
1	Digital Audiometer Set	
	(As per Specification)	



Dean
Govt. Medical College &
Hospital, Akola

DEPT. OF ENT
GOVT. MEDICAL COLLEGE, AKOLA

Specification of Diagnostic Audiometer SET.

1. It should have Air conduction, bone conduction, free field and speech audiometric with Built In Free Field Amplifier testing
 - Pure tone, pulse tone, warble tone, narrow band and White Noise
 - Special Tests such as SISI, Tone Decay, ABLB, Stenger, Luscher, Free field
2. Two separate channels with attenuators
3. Microprocessor controlled
4. Free field option
5. Online computer connection
6. Frequencies: Air conduction : 125-8000 Hz
Bone conduction : 250-8000 Hz
- 7. Air conduction**
Level range: -10 to 120+ dBHL, Headphone
- 8. Bone conduction**
Level range: -10 to max. 80+ dBHL, BC receiver head band
9. Speech – 10 to 100+ db HL
10. Masking – 10 to 100+ dBHL with facility for synchronized masking
11. Signal level should be electronically indicated
12. Large digital display should be present
13. Facility for the external inputs like CD player, Tape recorder, Microphone should be present
14. Should have built in computer interface
15. Should have facility for printing
16. Calibration should comply with latest IEC & ISO standard.
17. Talk back Mike
18. Carrying case
- 19. Free field**
Level range: -10 to max. 90 dBHL
Test signals: Pure tones, pulse and warble tone
Test frequencies: 125 – 12,000 Hz,
Level steps: 5 dB
Masking signals: narrowband noise, white noise, synchronized automation, speech noise masking

PTO


Prof. & Head,
Deptt. of E.N.T.
Medical College, Akola