

Re Quotation Calling

NO. PUR / Store / GMC / Akola/036 /21
Office of the Dean Govt. Medical College,
Akola 444001
Date : 4 / 2 / 2021

TO,

Subject : Submission of Quotation for **Drill Cum K-wire System (Battery & Power system)** to Dean Govt. Medical College, Akola .

Dear Sir,

- 1) The undersigned invited sealed quotations for the items mentioned below/on/reverse or as per enclosed statement on the following terms & conditions.
- 2) The Prices quoted should be free delivery to college & Hospital premises for local dealers and FOR Akola for outsiders.
- 3) The serial number of the items should not be changed while quoting rates. You may drop the item if not interested.
- 4) Rates should be quoted strictly for the item specified in the list and for standard quality of goods. In case of alternate offer, the detailed specifications, name of manufacturer or make etc must invariably be stated. Specifications other than specified in the schedule may be liable for rejection on even though lowest.
- 5) The quotation submitted will be valid for a period of ONE YEAR from the date of acceptance.
- 6) Delivery period should be stated specifically the like ready stock or two weeks or four weeks etc and should be firm and supply of stores if ordered should be made available within the stipulated period. Failures are liable for delist from further enquiries.
- 7) The quotation received after due date will not be accepted. The quotation should be submitted in sealed cover. Unsealed quotations which are not properly sealed will not be accepted. The last or receiving date of enquires is ... 1.2.21, 2.2.21, upto 5.00 P.M.
- 8) Supply of stores should be made in one installment unless otherwise ordered place meal supply will not be accepted. Payment will be made within 4 to 8 weeks after report of working etc of the material in full quantity and bill in quadruplicate under only satisfactory report of working etc. of the stores. Condition of advance payment through Bank, part payment etc. will not be accepted.
- 9) The price quoted is inclusive of all taxes, duties if payable like Customs/Excise/CST/ST the breakup of the taxes should also be shown separately where necessary. The sales tax and registration number should be quoted in your letter. Exemption of taxes on 'AF'/Octroi exemption from etc be separately stated.
- 10) Quotations if asked with samples if not accompanies with sample will be liable for rejection even they are lowest. Samples should be sent with a labels attached quoting our ref. No of enquiry and item no etc.
- 11) The Dean, Govt. Medical College & Hospital, Akola does not pledge himself to accept the lowest or any quotations and reserves to himself right of acceptance or rejection any of quotations which does not suit to his requirements.
- 12) **VERY IMPORTANCE :-** In case you are not interested in quoting your rates; in reply to this enquiry a line in reply is must; failing are liable for delisting their names for further enquiries from our list and no further request in this matter will be entertained.
- 13) Literature, Instruction manual, working leaflet etc may also be sent with quotation.

Sr.no	Name of Reagent	Qty
1	Drill Cum K-wire System (Battery & Power system) As Specification	

Note- Prices must be quoted inclusive of all taxes/charges.


Dean

Govt. Medical College Akola

Govt. Medical College Akola.

Specification of Drill System Cum Reamer & K-wire System

- Fully Autoclavable & cannulated.
- Battery Operated.
- Varying speed control with finger trigger
- Convenient fwd & Rev direction control with help of switch for drill hand piece.
- High speed drilling and high torque cutting for drill.
- 1200 RPM drilling speed.
- Autoclaved silicon wire, six feet long provided for electrical mains operation.
- Provision to fit standard Jacob Chuck/quick coupling chuck.

Battery Operated Dual Function Reamer System

- Fully Autoclavable & cannulated.
- Battery Operated.
- Varying speed control with finger trigger
- High speed drilling and high torque
- Provision to fit standard Jacob Chuck/quick coupling chuck.
- Works on Main as well as battery
- Can fit any standard Reamers.