

Quotation By Email

6010

NO. PUR / Store / GMC /Akola/ /18  
Office of the Dean Govt. Medical  
& Hosp. College, Akola 444001  
Date : /10/2018

29/11/18

TO,


Subject : Submission of Quotation for the Supply of Instrument etc.  
to Dean Govt. Medical College, Akola .

Dear Sir,

You are requested to submit your scaled quotation for the supply of Instrument/Machine to this office with the terms and conditions mentioned below

- 1) The quotation should be send in scaled cover (duly sealed wax) by the post or by hand delivery so as to reach this office on be 7/12/2018
- 2) Quotation will be open in the office of the dean on 10/12/2018
- 3) Quotation which are received late will not be accepted and any circumstances.
- 4) The words quotation for and last date , must be mentioned on envelope .
- 5) Rate should be for free delivery at the Dean , Govt. Medical college & Hos. Akola. premises only .
- 6) Quoted good should be strictly according to the make & the specification asked in enquiry, and make of item by you should be stated in your quotation .
- 7) In case the order are placed with you , the order will have to be executed in full within the stipulated time . If you fail to comply , the orders may be cancelled and you will be declared defaulter , no enquiry will be sent in future without assigning any reason.
- 8) The undersigned reserves the right to accept or reject any or all quotation. The quotation should be sent in the name of the / Dean , medical college Akola .
- 9) Please quote the sale tax /GST No registration Certificate, Pan Card., with your quotation.

Sr.no	Name of Reagent	Qty
1	Air Compressor As per Specification	

  
 Dean  
 Govt. Medical College &  
 Hospital, Akola

**Govt. Medical College, Akola**  
**Dept. of Dentistry**

**Specifications for Air Compressor for Dental Chairs**

Name of the Equipment	Specifications
Air Compressor for Dental Chairs	2 HP air compressor made of steel body, having the capacity to run three dental chairs. Compressor should be oil free, noise free. Technical Specifications: * Motor: Single Phase, 2 HP * RPM: 1440 * Operating Voltage: 210-240 volts * Maximum Pressure: 8 bar * Free air Delivery: <u>150Lpm@90 Psi</u> * Piston displacement: 240 Lpm * Tank Volume: 50 Liters * Auto on/off * Pump: No Lubrication required * Net Weight: 48 kgs



**(Dr. Dinesh Naitam)**  
**Signature of HOD**  
**Dept. of Dentistry**  
**GMC, Akola**